



TRUE VALUE CALIBRATION SERVICES PRIVATE LIMITED

NABL SCOPE

Mechanical- Weighing Scale and Balance

Sr.No.	Instrument	Range	CMC (±)
Permanent Facility			
1.	Electronic Weighing Balance (Class I) Readability. 0.001 mg	0to 5 g	0.004 mg
2.	Electronic Weighing Balance (Class I) Readability:1 mg	0to 40 kg	7 mg
3.	Electronic Weighing Balance (Class I) Readability: 0.001 mg	0 to 1 g	0.003 mg
4.	Electronic Weighing Balance (Class I) Readability: 0.01 mg	0 to 22 g	0.02 mg
5.	Electronic Weighing Balance & Mass Comparator (Class I) Readability: 0.1 mg	0 to 5100 g	0.6 mg
6.	Electronic Weighing Balance (Class I) Readability.: 0.001 mg	0 to 2 g	0.004 mg
7.	Electronic Weighing Balance (Class I) Readability: 0.01 mg	0 to 220 g	0.028 mg
8.	Electronic Weighing Balance (Class II) Readability 0.1 g	0 to 64 kg	0.5 g
Site Facility			
1.	Electronic Weighing Balance (Class I) Readability: 0.001 mg	0 to 11 g	0.004 mg
2.	Electronic Weighing Balance (Class I) Readability: 0.1 mg	0 to 5100 g	0.6 mg
3.	Electronic Weighing Balance (Class I) Readability: 1 mg	0 to 40 kg	7 mg
4.	Electronic Weighing Balance (Class I) Readability: 0.01 mg	0 to 22 g	0.02 mg
5.	Electronic Weighing Balance (Class III) Readability 200 g	0 to 1000 kg	191 g
6.	Electronic Weighing Balance (Class III) Readability 100 g	0 to 500 kg	82 g
7.	Electronic Micro Balance (Class I) Readability: 0.001 mg	0to 1 g	0.003 mg
8.	Electronic Weighing Balance(Class I) Readability: 0.001 mg	0 to 2 g	0.004 mg
9.	Electronic Weighing Balance (Class I) Readability: 0.01 mg	0 to 220 g	0.028 mg
10.	Electronic Weighing Balance (Class III) Readability: 10 g	0 to 300 kg	9.34 g
11	Electronic Weighing Balance (Class III) Readability: 5g	0 to 100 kg	5.78 g
12.	Electronic Weighing Balance (Class I) Readability: 0.001 mg	0 to 5 g	0.004 mg
13.	Electronic Weighing Balance (Class II) Readability: 0.1 g	0 to 64 kg	0.5 g
14.	Electronic Weighing Balance (Class III) Readability: 500 g	0 to 5000 kg	460.8 g